

Test Code: M_DW_ICP_ET
Test Number: E200.7
Test Name: EPA 600 4.1.4, 200.7R4.4, Metals by ICP (Total)
Matrix: Drinking Water **Units:** mg/L

**METHOD DETECTION/REPORTING
LIMITS**

Type	Analyte	Synonym	MDL	PQL	EPA Lbl
A	Aluminum		0.015	0.05	NELAP
A	Barium		0.0024	0.005	NELAP
A	Beryllium		0.0003	0.001	NELAP
A	Boron		0.0055	0.02	
A	Cadmium		0.0003	0.002	NELAP
A	Calcium		0.019	0.05	NELAP
A	Chromium		0.004	0.01	NELAP
A	Copper		0.003	0.01	NELAP
A	Iron		0.007	0.02	NELAP
A	Magnesium		0.0051	0.01	NELAP
A	Manganese		0.0016	0.005	NELAP
A	Molybdenum		0.0032	0.01	
A	Nickel		0.0033	0.01	NELAP
A	Potassium		0.024	0.1	
A	Silver		0.003	0.01	NELAP
A	Sodium		0.024	0.05	NELAP
A	Vanadium		0.0032	0.01	
A	Zinc		0.0021	0.01	NELAP