

---

**Test Code:** M\_DW\_GF\_MT  
**Test Number:** M3113B  
**Test Name:** Standard Methods 18th Ed. 3030 E, 3113 B, Metals by GFAA  
**Matrix:** Drinking Water     **Units:** mg/L

---

**METHOD DETECTION/REPORTING  
LIMITS**

---

<b>Type</b>	<b>Analyte</b>	<b>Synonym</b>	<b>MDL</b>	<b>PQL</b>	<b>EPA Lbl</b>
A	Antimony		0.0017	0.005	NELAP
A	Arsenic		0.0007	0.003	NELAP
A	Lead		0.0004	0.002	NELAP
A	Selenium		0.002	0.006	NELAP