

Test Code: M_DW_GF_TL_200.9

**METHOD DETECTION/REPORTING
LIMITS**

Test Number: E200.9

Test Name: EPA 600 4.1.1, EPA 200.9 R2.2 Metals by GFAA

Matrix: Drinking Water Units: mg/L

Type	Analyte	Synonym	MDL	PQL	EPA Lbl
A	Thallium		0.0005	0.002	NELAP
