

Test Code: M_DW_ICPMS_ET
Test Number: E200.8
Test Name: EPA 600 4.1.4, 200.8 R5.4, Metals by ICPMS (Total)
Matrix: Drinking Water **Units:** mg/L

METHOD DETECTION/REPORTING LIMITS

Type	Analyte	Synonym	MDL	PQL	EPA Lbl
A	Antimony		0.00011	0.001	NELAP
A	Arsenic		0.00037	0.001	NELAP
A	Barium		0.00005	0.0002	NELAP
A	Beryllium		0.00002	0.0002	NELAP
A	Cadmium		0.00002	0.0002	NELAP
A	Chromium		0.00023	0.002	NELAP
A	Copper		0.00006	0.0002	NELAP
A	Lead		0.00002	0.0002	NELAP
A	Manganese		0.00015	0.001	NELAP
A	Molybdenum		0.00002	0.0002	NELAP
A	Nickel		0.00003	0.0002	NELAP
A	Selenium		0.00008	0.0002	NELAP
A	Silver		0.00002	0.0002	NELAP
A	Thallium		0.00004	0.0002	NELAP
A	Zinc		0.00025	0.002	NELAP
X	Aluminum		0.00101	0.005	*