

Test Code: SV_8151S_O

METHOD DETECTION/REPORTING
LIMITS

Test Number: SW8151A

Test Name: SW-846 3580A, 8151A Chlorinated Herbicides by GC/ECD

Matrix: Waste

Units: mg/Kg

Type	Analyte	Synonym	MDL	PQL	EPA Lbl
A	2,4,5-T		0.066	0.2	*
A	2,4,5-TP (Silvex)		0.062	0.2	*
A	2,4-D		0.22	0.4	*
A	2,4-DB		0.043	0.2	*
A	3,5-Dichlorobenzoic acid		0.056	0.2	*
A	4-Nitrophenol		0.094	0.2	*
A	Acifluorfen		0.034	0.2	*
A	Bentazon		0.061	0.2	*
A	Chloramben		0.055	0.2	*
A	Dalapon		0.235	1.3	*
A	DCPA	DCPA	0.11	0.2	*
A	Dicamba		0.052	0.2	*
A	Dichlorprop	Dichloroprop	0.05	0.2	*
A	Dinoseb		0.035	0.2	*
A	MCPA		2.73	45	*
A	MCPP		8.28	30	*
A	Pentachlorophenol		0.039	0.1	*
A	Picloram		0.052	0.2	*