

Test Code: SV_608E_W

METHOD DETECTION/REPORTING
LIMITS

Test Number: EPA608

Test Name: EPA 600 608, Chlorinated Pesticides & PCBs by GC/ECD

Matrix: Aqueous

Units: µg/L

Type	Analyte	Synonym	MDL	PQL	EPA Lbl
A	4,4'-DDD		0.012	0.05	NELAP
A	4,4'-DDE		0.015	0.05	NELAP
A	4,4'-DDT		0.014	0.05	NELAP
A	Aldrin		0.013	0.05	NELAP
A	alpha-BHC		0.022	0.05	NELAP
A	alpha-Chlordane		0.013	0.05	*
A	Aroclor 1016		0.2	1	NELAP
A	Aroclor 1221		0.16	1	NELAP
A	Aroclor 1232		0.12	1	NELAP
A	Aroclor 1242		0.13	1	NELAP
A	Aroclor 1248		0.16	1	NELAP
A	Aroclor 1254		0.22	1	NELAP
A	Aroclor 1260		0.1	1	NELAP
A	beta-BHC		0.032	0.05	NELAP
A	delta-BHC		0.02	0.05	NELAP
A	Dieldrin		0.015	0.05	NELAP
A	Endosulfan I		0.021	0.05	NELAP
A	Endosulfan II		0.014	0.05	NELAP
A	Endosulfan sulfate		0.014	0.05	NELAP
A	Endrin		0.015	0.05	NELAP
A	Endrin aldehyde		0.022	0.05	NELAP
A	Endrin Ketone		0.015	0.05	*
A	gamma-BHC (Lindane)		0.006	0.05	NELAP
A	gamma-Chlordane		0.016	0.05	*
A	Heptachlor		0.024	0.05	NELAP
A	Heptachlor epoxide		0.014	0.05	NELAP
A	Methoxychlor		0.022	0.05	NELAP
A	Toxaphene		0.087	0.5	NELAP
C	Chlordane		0.021	0.1	NELAP
C	Polychlorinated Biphenyls (Total)		1.09	7	NELAP