

Test Code: SV_608.3E_W

METHOD DETECTION/REPORTING
LIMITS

Test Number: EPA608

Test Name: EPA 600 608.3, Chlorinated Pesticides & PCBs by GC/ECD

Matrix: Aqueous

Units: µg/L

Type	Analyte	Synonym	MDL	PQL	EPA Lbl
A	4,4'-DDD		0.011	0.033	*
A	4,4'-DDE		0.004	0.012	*
A	4,4'-DDT		0.012	0.036	*
A	Aldrin		0.004	0.012	*
A	alpha-BHC		0.003	0.009	*
A	alpha-Chlordane		0.014	0.042	*
A	Aroclor 1016		0.065	0.095	*
A	Aroclor 1221		0.065	0.095	*
A	Aroclor 1232		0.065	0.095	*
A	Aroclor 1242		0.065	0.095	*
A	Aroclor 1248		0.065	0.095	*
A	Aroclor 1254		0.065	0.095	*
A	Aroclor 1260		0.065	0.095	*
A	beta-BHC		0.006	0.018	*
A	delta-BHC		0.009	0.027	*
A	Dieldrin		0.002	0.006	*
A	Endosulfan I		0.014	0.042	*
A	Endosulfan II		0.004	0.012	*
A	Endosulfan sulfate		0.066	0.198	*
A	Endrin		0.006	0.018	*
A	Endrin aldehyde		0.023	0.07	*
A	Endrin Ketone		0.015	0.07	*
A	gamma-BHC (Lindane)		0.004	0.012	*
A	gamma-Chlordane		0.014	0.042	*
A	Heptachlor		0.003	0.009	*
A	Heptachlor epoxide		0.083	0.249	*
A	Methoxychlor		0.022	0.042	*
A	Toxaphene		0.24	0.72	*
C	Chlordane		0.028	0.084	*
C	Polychlorinated Biphenyls (Total)		0.455	0.665	*