

Test Code: SV_615E_W
Test Number: EPA615
Test Name: EPA 600 615, Chlorinated Herbicides by GC
Matrix: Aqueous **Units:** µg/L

METHOD DETECTION/REPORTING LIMITS

| Type | Analyte | Synonym | MDL | PQL | EPA Lbl |
|------|--------------------------|---------|------|-----|---------|
| A | 2,4,5-T | | 0.05 | 0.2 | NELAP |
| A | 2,4,5-TP (Silvex) | | 0.04 | 0.2 | NELAP |
| A | 2,4-D | | 0.04 | 0.2 | NELAP |
| A | 2,4-DB | | 0.05 | 0.2 | * |
| A | 3,5-Dichlorobenzoic Acid | | 0.06 | 0.2 | * |
| A | 4-Nitrophenol | | 0.21 | 0.3 | * |
| A | Acifluorfen | | 0.02 | 0.2 | * |
| A | Bentazon | | 0.06 | 0.2 | * |
| A | Chloramben | | 0.04 | 0.2 | * |
| A | Dalapon | | 0.16 | 1.3 | * |
| A | DCPA | | 0.05 | 0.2 | * |
| A | Dicamba | | 0.04 | 0.2 | NELAP |
| A | Dichlorprop | | 0.05 | 0.2 | * |
| A | Dinoseb | | 0.02 | 0.2 | * |
| A | MCPA | | 7.55 | 45 | * |
| A | MCPP | | 7.77 | 30 | * |
| A | Pentachlorophenol | | 0.03 | 0.1 | * |
| A | Picloram | | 0.05 | 0.2 | * |