

**Test Code:** SV\_548E\_W  
**Test Number:** EPA548.1  
**Test Name:** EPA 600 548.1, Endothall by GC/MS  
**Matrix:** Drinking Water    **Units:** µg/L

**METHOD DETECTION/REPORTING  
LIMITS**

Type	Analyte	Synonym	MDL	PQL	EPA Lbl
A	Endothall		0.21	1	*